



Republic of the Philippines
 Department of Agriculture
BUREAU OF ANIMAL INDUSTRY
 5 Visayas Avenue, Barangay Vasra, Quezon City 1128

(+632)8528-2240 director@bai.gov.ph www.bai.gov.ph @bai.gov.ph

June 05, 2024

SUPPLEMENTAL / BID BULLETIN
No. 2024-008

This Supplemental/Bid Bulletin No. 2024-008 is issued by the BAI Bids and Awards Committee to all participating bidders to clarify, amend and/or modify certain provisions and requirements for the procurement of various projects.

All portions of the Bidding Documents affected by these amendments shall be made to conform the same.

Amendments/inclusions/clarifications made herein shall be considered as integral part of the Bidding Documents.

FROM	TO
IB -2024-024 SUPPLY AND DELIVERY OF CONTAINER VAN ABC - P6,400,000.00	
CONTAINER VAN 40-footer, Fabricated, Knock-down unit assembled at site D. Flooring <ul style="list-style-type: none"> ¼ Marine Plywood sub-floor on wooden container floor 2 mm Vinyl Planks E. Painting <ul style="list-style-type: none"> Enamel painted interior and exterior (Boysen or equivalent) 	CONTAINER VAN (shipping container) 40-footer, Fabricated, Knock-down unit assembled at site Height: 9.6 ft Corrugated Panel thickness: 2mm D. Flooring <ul style="list-style-type: none"> ¼ Marine Plywood sub-floor on wooden container floor 2 mm Vinyl Planks treated flooring materials E. Painting <ul style="list-style-type: none"> Enamel painted interior and exterior (Boysen or equivalent) - 3 coats for interior, 2 coats for exterior Inclusion: excavation and hauling of unsuitable materials Bidder must attach perspective and vicinity map as their brochure Bidder must conduct site inspection prior to the submission of bid proposal Bidder must conduct site inspection prior to the submission of bid proposal



"Our organization is certified according to ISO 9001"



Republic of the Philippines
 Department of Agriculture
BUREAU OF ANIMAL INDUSTRY
 5 Visayas Avenue, Barangay Vasra, Quezon City 1128

(+632)8528-2240 director@bai.gov.ph www.bai.gov.ph @bai.gov.ph

IB -2024-025 SUPPLY AND DELIVERY OF BIOSAFETY CABINET ABC – P7,000,000.00	
<p>LOT I: BIOSAFETY CABINET Class II A2 – 3 ft.</p> <p>Specifications:</p> <p>Filters: HEPA H14 filters according to EN 1822: Efficiency of 99.995%</p> <p>Front Entry AirSpeed /Extraction Flow Rate: 0.55 m/s/295 m³/h</p> <p>Laminar Flow Speed/Laminar Flow Rate: 0.35m/s/669m³/h</p> <p>Voltage: 230 V</p> <p>Warranty: IQ, OQ, PQ – certificates after installation</p> <p>Certificate of Satisfactory Performance from previous installation</p> <p>Inclusions:</p> <p>With laboratory chair in each unit</p> <ul style="list-style-type: none"> - Weight capacity: 200 kg <p>Time of delivery: 60 days</p> <p style="text-align: center;">*****</p> <p>LOT II: BIOSAFETY CABINET Class II A2 – 4 ft.</p> <p>Specifications:</p> <p>Internal dimensions (WxDxH): 1178 x 654 x 724 mm</p> <p>External dimensions (WxDxH): 1362 x 817 x 1546 mm</p> <p>Downflow: 0.30 m/s Inflow: 0.53 m/s Exhaust Filter: 11 1/2-inch (292 mm) 99.995% Efficiency HEPA Filter Seal Type: HEPA Filter</p> <p>Warranty: IQ, OQ, PQ – certificates after installation</p>	<p>Filters: HEPA H14 filter or higher according to EN 1822: Efficiency of 99.995%</p> <p>Front Entry AirSpeed /Extraction Flow Rate: 0.53 -0.55 m/s/295 m³/h</p> <p>Laminar Flow Speed/Laminar Flow Rate: 0.33 - 0.35m/s/669m³/h</p> <p>Voltage: 220-230 V</p> <p>IQ, OQ, PQ – certificates after installation/ISO certification from the manufacturer</p> <p>Certificate of Satisfactory Performance from previous installation (same brand of BSC)</p> <ul style="list-style-type: none"> - Weight capacity: 100-120 kg <p>Time of delivery: 90 days</p> <p style="text-align: center;">*****</p> <p>Maximum Internal dimensions (WxDxH): 1259 x 654 x 724 mm</p> <p>Maximum External dimensions (WxDxH): 1362 x 817 x 1546 mm Downflow: 0.30- 0.35 m/s Inflow: 0.53 -0.55 m/s Exhaust Filter: 11 1/2-inch (292 mm) 99.995% Efficiency or better efficiency level HEPA Filter Seal Type: HEPA Filter or better</p> <p>IQ, OQ, PQ – certificates after installation /ISO certification from the manufacturer</p>



"Our organization is certified according to ISO 9001"



Republic of the Philippines
 Department of Agriculture
BUREAU OF ANIMAL INDUSTRY
 5 Visayas Avenue, Barangay Vasra, Quezon City 1128

(+632)8528-2240 director@bai.gov.ph www.bai.gov.ph @bai.gov.ph

Certificate of Satisfactory Performance from previous installation Inclusions: 3000 kVA AVR compatible with the unit With laboratory chair in each unit - Weight capacity: 200 kg Time of delivery: 60 days	Certificate of Satisfactory Performance from previous installation (same brand of BSC) 3 kVA AVR compatible with the unit - Weight capacity: 100-120 kg Time of delivery: 90 days
IB -2024-026 SUPPLY AND DELIVERY OF QPCR CALIBRATION KITS ABC - P1,758,816.00	
Expiry Date: At least 1 year upon delivery. Delivery Period: 60 calendar days	At least 8 months upon delivery. Delivery Period: 45 calendar days
IB -2024-027 SUPPLY AND DELIVERY OF OPTICAL PLATES AND FILMS ABC - P16,622,000.00	
Delivery Period: 60 calendar days	Delivery Period: 90 calendar days

Please be guided accordingly.


CHERYL ROSE B. CAYAD-AN, DVM, MDM
 Chairperson
 Bids and Awards Committee 



"Our organization is certified according to ISO 9001"