



Republic of the Philippines  
 Department of Agriculture  
**BUREAU OF ANIMAL INDUSTRY**  
 5 Visayas Avenue, Barangay Vasra, Quezon City 1128

(+632)8528-2240 director@bai.gov.ph www.bai.gov.ph @bai.gov.ph

June 05, 2024

**SUPPLEMENTAL / BID BULLETIN**  
**No. 2024-008**

This Supplemental/Bid Bulletin No. 2024-008 is issued by the BAI Bids and Awards Committee to all participating bidders to clarify, amend and/or modify certain provisions and requirements for the procurement of various projects.

All portions of the Bidding Documents affected by these amendments shall be made to conform the same.

Amendments/inclusions/clarifications made herein shall be considered as integral part of the Bidding Documents.

FROM	TO
<b>IB -2024-024</b> <b>SUPPLY AND DELIVERY OF CONTAINER VAN</b> <b>ABC - P6,400,000.00</b>	
<b>CONTAINER VAN</b> 40-footer, Fabricated, Knock-down unit assembled at site  D. Flooring <ul style="list-style-type: none"> <li>¼ Marine Plywood sub-floor on wooden container floor</li> <li>2 mm Vinyl Planks</li> </ul> E. Painting <ul style="list-style-type: none"> <li>Enamel painted interior and exterior (Boysen or equivalent)</li> </ul>	<b>CONTAINER VAN (shipping container)</b> 40-footer, Fabricated, Knock-down unit assembled at site  <b>Height: 9.6 ft</b> <b>Corrugated Panel thickness: 2mm</b>  D. Flooring <ul style="list-style-type: none"> <li>¼ Marine Plywood sub-floor on wooden container floor</li> <li>2 mm Vinyl Planks</li> <li><b>treated flooring materials</b></li> </ul> E. Painting <ul style="list-style-type: none"> <li>Enamel painted interior and exterior (Boysen or equivalent)</li> <li>- <b>3 coats for interior, 2 coats for exterior</b></li> </ul> <b>Inclusion: excavation and hauling of unsuitable materials</b>  <b>Bidder must attach perspective and vicinity map as their brochure</b>  <b>Bidder must conduct site inspection prior to the submission of bid proposal</b>  <b>Bidder must conduct site inspection prior to the submission of bid proposal</b>



"Our organization is certified according to ISO 9001"



<b>IB -2024-025</b> <b>SUPPLY AND DELIVERY OF BIOSAFETY CABINET</b> <b>ABC – P7,000,000.00</b>	
<p><b>LOT I: BIOSAFETY CABINET Class II A2 – 3 ft.</b></p> <p><b>Specifications:</b></p> <p>Filters: HEPA H14 filters according to EN 1822: Efficiency of 99.995%</p> <p>Front Entry AirSpeed /Extraction Flow Rate: 0.55 m/s/295 m<sup>3</sup>/h</p> <p>Laminar Flow Speed/Laminar Flow Rate: 0.35m/s/669m<sup>3</sup>/h</p> <p>Voltage: 230 V</p> <p><b>Warranty:</b> IQ, OQ, PQ – certificates after installation</p> <p>Certificate of Satisfactory Performance from previous installation</p> <p><b>Inclusions:</b></p> <p>3000 kVA AVR compatible with the unit</p> <p><b>With laboratory chair in each unit</b></p> <ul style="list-style-type: none"> <li>- Weight capacity: 200 kg</li> </ul> <p>Time of delivery: 60 days</p> <p style="text-align: center;">*****</p> <p><b>LOT II: BIOSAFETY CABINET Class II A2 – 4 ft.</b></p> <p><b>Specifications:</b></p> <p>Internal dimensions (WxDxH): 1178 x 654 x 724 mm</p> <p>External dimensions (WxDxH): 1362 x 817 x 1546 mm</p> <p>Downflow: 0.30 m/s Inflow: 0.53 m/s Exhaust Filter: 11 1/2-inch (292 mm) 99.995% Efficiency HEPA Filter Seal Type: HEPA Filter</p> <p><b>Warranty:</b> IQ, OQ, PQ – certificates after installation</p>	<p>Filters: HEPA H14 filter <b>or higher</b> according to EN 1822: Efficiency of 99.995%</p> <p>Front Entry AirSpeed /Extraction Flow Rate: <b>0.53 -0.55 m/s/295 m<sup>3</sup>/h</b></p> <p>Laminar Flow Speed/Laminar Flow Rate: <b>0.33 - 0.35m/s/669m<sup>3</sup>/h</b></p> <p><b>Voltage: 220-230 V</b></p> <p><b>IQ, OQ, PQ – certificates after installation/ISO certification from the manufacturer</b></p> <p>Certificate of Satisfactory Performance from previous installation (<b>same brand of BSC</b>)</p> <p><b>3 kVA AVR compatible with the unit</b></p> <ul style="list-style-type: none"> <li>- Weight capacity: <b>100-120 kg</b></li> </ul> <p>Time of delivery: <b>90 days</b></p> <p style="text-align: center;">*****</p> <p><b>Maximum</b> Internal dimensions (WxDxH): <b>1259</b> x 654 x 724 mm</p> <p><b>Maximum</b> External dimensions (WxDxH): 1362 x 817 x 1546 mm</p> <p>Downflow: 0.30- <b>0.35 m/s</b> Inflow: <b>0.53 -0.55 m/s</b> Exhaust Filter: 11 1/2-inch (292 mm) 99.995% Efficiency <b>or better efficiency level</b> HEPA Filter Seal Type: HEPA Filter <b>or better</b></p> <p><b>IQ, OQ, PQ – certificates after installation /ISO certification from the manufacturer</b></p>





Republic of the Philippines  
 Department of Agriculture  
**BUREAU OF ANIMAL INDUSTRY**  
 5 Visayas Avenue, Barangay Vasra, Quezon City 1128

(+632)8528-2240 director@bai.gov.ph www.bai.gov.ph @bai.gov.ph

Certificate of Satisfactory Performance from previous installation  <b>Inclusions:</b> 3000 kVA AVR compatible with the unit  <b>With laboratory chair in each unit</b> - Weight capacity: 200 kg  Time of delivery: 60 days	Certificate of Satisfactory Performance from previous installation <b>(same brand of BSC)</b>  3 kVA AVR compatible with the unit  - Weight capacity: <b>100-120 kg</b>  Time of delivery: <b>90 days</b>
<b>IB -2024-026</b> <b>SUPPLY AND DELIVERY OF QPCR CALIBRATION KITS</b> <b>ABC - P1,758,816.00</b>	
Expiry Date: At least 1 year upon delivery.  Delivery Period: 60 calendar days	At least <b>8 months</b> upon delivery.  Delivery Period: <b>45</b> calendar days
<b>IB -2024-027</b> <b>SUPPLY AND DELIVERY OF OPTICAL PLATES AND FILMS</b> <b>ABC - P16,622,000.00</b>	
<b>Delivery Period: 60</b> calendar days	<b>Delivery Period: 90</b> calendar days

Please be guided accordingly.

  
**CHERYL ROSE B. CAYAD-AN, DVM, MDM**  
 Chairperson  
 Bids and Awards Committee 



"Our organization is certified according to ISO 9001"

*Masaganang Agrikultura, Maunlad na Ekonomiya*